



STATE OF WASHINGTON  
**DEPARTMENT OF ECOLOGY**  
PO Box 488 • Manchester, WA 98353-0488 • (360) 871-8840

July 22, 2025

Munaf Khan  
NVL Laboratories, Inc  
4708 Aurora Ave N  
Seattle, WA 98103

Dear Munaf Khan:

Thank you for your application for renewal in the Environmental Laboratory Accreditation Program. Attached is a new Certificate of Accreditation covering the one-year period beginning retroactively 7/18/2025, and a current Scope of Accreditation.

**Two acceptable PT results and a corrective action report (CAR) are required prior to the next renewal for the following parameter due to a PT failure in the past accreditation year:**

- *Arsenic by EPA 6010D\_ (7/18) in Solid and Chemical Materials*

Renewal of accreditation is based in part on review of your lab's performance over the past year as evidenced by participation in proficiency testing (PT) studies. In general, full accreditation is awarded for those parameters for which the two most recent PT results, if applicable, were rated satisfactory. Provisional accreditation is awarded if the latest of the two most recent PT results was rated "Not Acceptable" or only one PT result was submitted during the past twelve months. Accreditation is withheld for those parameters for which the two most recent PT results were rated "Not Acceptable" or no PT results were submitted during the past twelve months.

As a reminder, continued participation in the Ecology Lab Accreditation Program requires the lab to:

- Submit a renewal application and fees annually
- Report significant changes in facility, personnel, analytical methods, equipment, the lab's quality assurance (QA) manual or QA procedures as they occur
- **Participate in proficiency testing studies semi-annually, with the following exception: For each parameter where all PT results were satisfactory, you are required to submit only one PT result over this next year, and in subsequent years, as long as the results are satisfactory.**
- **Submit copies of current third-party Scopes of Accreditation when they are available.**

## Your Right To Appeal

You have a right to appeal Ecology's decision to the Pollution Control Hearing Board (PCHB) within 30 days of the date of receipt of this decision letter. The appeal process is governed by Chapter 43.21B RCW and Chapter 371-08 WAC. "Date of receipt" is defined in RCW 43.21B.001(2).

To appeal you must do the following within 30 days of the date of receipt of this decision:

- File your appeal and a copy of this decision with the PCHB (see addresses below). Filing means actual receipt by the PCHB during regular business hours.
- Serve a copy of your appeal and this decision on Ecology in paper form - by mail or in person. (See addresses below.) E-mail is not accepted.

You must also comply with other applicable requirements in Chapter 43.21B RCW and Chapter 371-08 WAC.

## Address And Location Information

### Street Addresses:

#### Department of Ecology

Attn: Appeals Processing Desk  
300 Desmond Drive SE  
Lacey, WA 98503

#### Pollution Control Hearings Board

1111 Israel RD SW  
STE 301  
Tumwater, WA 98501

### Mailing Addresses:

#### Department of Ecology

Attn: Appeals Processing Desk  
PO Box 47608  
Olympia, WA 98504-7608

#### Pollution Control Hearings Board

PO Box 40903  
Olympia, WA 98504-0903

### E-Mail Address:

#### Department of Ecology

Not currently available (see WAC 371-08)

#### Pollution Control Hearings Board

Pchb-shbappeals@elaho.wa.gov

If you have any questions concerning the accreditation of your lab, please contact Julia Dolan at (360) 871-8842 or by e-mail at [Jdol461@ECY.WA.GOV](mailto:Jdol461@ECY.WA.GOV), or Andrea Allen at (564) 233- 5740 or by e-mail at [abid461@ECY.WA.GOV](mailto:abid461@ECY.WA.GOV).

Sincerely,

A handwritten signature in black ink that reads "Rebecca Wood". The signature is written in a cursive style with a large, sweeping initial 'R' and a long, horizontal flourish at the end.

Rebecca Wood  
Lab Accreditation Unit Supervisor

The State of  
Department



Washington  
of Ecology

**NVL Laboratories, Inc**

**Seattle, WA**

has complied with provisions set forth in Chapter 173-50 WAC and is hereby recognized by the Department of Ecology as an ACCREDITED LABORATORY for the analytical parameters listed on the accompanying Scope of Accreditation.

This certificate is effective 7/18/2025 and shall expire 7/17/2026.

Witnessed under my hand on 7/22/2025.

Rebecca Wood  
Lab Accreditation Unit Supervisor

Laboratory ID  
C797

# WASHINGTON STATE DEPARTMENT OF ECOLOGY

## ENVIRONMENTAL LABORATORY ACCREDITATION PROGRAM

### SCOPE OF ACCREDITATION

**NVL Laboratories, Inc**

**Seattle, WA**

is accredited for the analytes listed below using the methods indicated. Full accreditation is granted unless stated otherwise in a note. EPA is the U.S. Environmental Protection Agency. SM is "Standard Methods for the Examination of Water and Wastewater." SM refers to EPA approved method versions. ASTM is the American Society for Testing and Materials. USGS is the U.S. Geological Survey. AOAC is the Association of Official Analytical Chemists. Other references are described in notes.

<b>Matrix/ Analyte</b>	<b>Method</b>	<b>Note</b>
<b>Drinking Water</b>		
Copper	EPA 200.8 5.4 1994	
Lead	EPA 200.8 5.4 1994	
Copper	EPA 200.9 Rev 2.2 (1994)	
Lead	EPA 200.9 Rev 2.2 (1994)	
<b>Solid and Chemical Materials</b>		
Arsenic	EPA 6010D (7/18)	
Barium	EPA 6010D (7/18)	
Cadmium	EPA 6010D (7/18)	
Chromium	EPA 6010D (7/18)	
Copper	EPA 6010D (7/18)	
Lead	EPA 6010D (7/18)	
Nickel	EPA 6010D (7/18)	
Selenium	EPA 6010D (7/18)	
Silver	EPA 6010D (7/18)	
Zinc	EPA 6010D (7/18)	
Mercury	EPA 7471B (2/07)	
Aroclor-1016 (PCB-1016)	EPA 8082A (2/07)	
Aroclor-1221 (PCB-1221)	EPA 8082A (2/07)	
Aroclor-1232 (PCB-1232)	EPA 8082A (2/07)	
Aroclor-1242 (PCB-1242)	EPA 8082A (2/07)	
Aroclor-1248 (PCB-1248)	EPA 8082A (2/07)	
Aroclor-1254 (PCB-1254)	EPA 8082A (2/07)	
Aroclor-1260 (PCB-1260)	EPA 8082A (2/07)	
Asbestos	EPA 600/R-93-116	1

### Accredited Parameter Note Detail

(1) Accreditation based in part on recognition of US Department of Commerce NIST NVLAP accreditation.



07/22/2025

---

Authentication Signature  
Rebecca Wood, Lab Accreditation Unit Supervisor

---

**Date**